

IMMEUBLE AURELIUM - 1 COURS DE L'ILE SEGUIN 92100 BOULOGNE-BILLANCOURT – FRANCE

TEL: +33 (0)1 48 10 62 62 FAX: +33 (0)1 48 91 22 99

TECHNICAL DATA SHEET

Page 1/1

PRODUCT NAME - DESIGNATION

VITALITEC® HEMOSTATIC CLIP APPLIERS

REFERENCE(S): various references depending on size and jaw design.

INDICATIONS

VITALITEC® Appliers are intended for the loading, holding and optimal closure of VITALITEC® hemostatic clips. These appliers must be used exclusively for SLS-CLIP® and CLIP 9 clips.

COMPOSITION

Applier: Stainless Steel **Ring coating**: Polyamid 11

CHARACTERISTICS

The coating color code identifies the applier adapted to the size of clip used.

Clip size	Color code
Micro	White
Small	Yellow
Small	Red
Small/Medium	Lilac
Medium	Blue
Medium/Large	Green
Large	Orange

Applier length: 15 cm, 20 cm, 28 cm

Jaw angulation : 25° , 65°

Jaw design: standard, fine and lightweight, extra-fine and lightweight

PACKAGING

Order Unit: Bag containing one clip applier.

STERILIZATION / EXPIRY DATE

Product delivered non-sterile – $\underline{\text{To be sterilized before use}}$ - $\underline{\text{Reusable}}$ devices

The clip appliers must be steam autoclaved in compliance with local regulatory requirements.

The Instructions For Use describe the sterilization method validated for these instruments.

CLEANING / DECONTAMINATION

The clip appliers must be cleaned and disinfected in compliance with local regulatory requirements. The Instructions For Use describe the validated cleaning and disinfection methods for these instruments.

CONTRAINDICATIONS

These appliers are not meant for any other use than indicated.

Never use appliers unsuited to the clip size.

Never use an applier which appears to be defective.

Never try to repair an instrument yourself.

CE CLASSIFICATION

Class I

French manufacturer

CE

Peters Surgical is certified ISO 9001, ISO 13485 and ISO 11135-1 and the products are made according to the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.